



PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

OCM Pro CF (incorporating POA Sensor) Echo Profile Flowmeter

manufactured by:

NIVUS GmbH

Im Täle
75031 Eppingen
Germany

has been assessed by Sira Certification Service
and for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Water Monitoring
Equipment Part 3, Version 2.1 dated March 2008**

Certification Ranges :

Velocity :	0 to 2 m/s
Depth :	0.1 to 1 m

Project No:	16W22558
Certificate No:	Sira MC110189/00
Initial Certification:	24 June 2011
This Certificate Issued	24 June 2011
Renewal Date:	23 June 2016

Technical Director

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service

12 Acorn Industrial Park, Crayford Road, Crayford
Dartford, Kent, UK, DA1 4AL
Tel: 01322 520500 Fax: 01322 520501

PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

OCM Pro CF (incorporating POA pipe sensor) Ultrasonic Echo Profile Flowmeter

Manufactured by:

NIVUS GmbH

Im Täle
75031 Eppingen
Germany

Has been assessed by Sira Certification Service
And for the conditions stated on this certificate complies with:

MCERTS Performance Standards for Water Monitoring Equipment Part 3, Version 2.3 dated March 2008

Certification Ranges :

Velocity : 0.05 to 2.5 m/s
Fluid depth : 0.075 to 0.6 m

Project No. : 16W28270
Certificate No : Sira MC130219/00
Initial Certification : 31 January 2013
This Certificate issued : 31 January 2013
Renewal Date : 30 January 2018

R Cooper | Eng MInst MC
Technical Director

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service

12 Acorn Industrial Park, Crayford Road, Crayford
Dartford, Kent, UK DA1 4AL
Tel: +44 (0)1322 520500 Fax: +44 (0)1322 520501



This certificate may only be reproduced in its entirety and without change

Registered Office: Rake Lane, Eccleston, Chester, UK CH4 9JN

To authenticate the validity of this certificate please visit www.siracertification.com/mcerts